

# Neonatology

- [Infant Warmer and Resuscitation Table](#)
- [Oxygen Concentrator](#)
- [Oxygen Concentrator](#)
- [Oxygen Concentrator](#)
- [Oxygen Concentrator](#)
- [Infusion Pump](#)
- [CPAP Maschine](#)
- [Electric Suction](#)

# Infant Warmer and Resuscitation Table

Operational

Last updated 2026-05-21

Current location Neonatology

## Core details

<b>Asset ID</b>	MT-60	<b>Department</b>	Neonatology
<b>Device type</b>	Infant Warmer and Resuscitation Table	<b>Location</b>	Neonatology
<b>Manufacturer</b>	Medicor Elektronika	<b>Model</b>	BLR-2100
<b>Serial number</b>	BDY00008	<b>Year</b>	2019
<b>Responsible</b>	—	<b>Ownership</b>	Donation

## Technical

<b>Power</b>	—	<b>Connectivity</b>	—
<b>Other</b>	—		
<b>Software</b>	—		
<b>Consumables</b>	—		

## Maintenance log

PM = Preventive maintenance, CM = Corrective maintenance

Date	Type	Action	By	Ref/Ticket	Next due
— no entries yet —					

## Notes

Known issues

—

Spare parts on-site

—

Vendor / technician contact

Medicor Elektronika Zrt. Budaapest Hungary

Escalation

—

# Oxygen Concentrator

Operational

Last updated 2026-05-21

Current location Neonatology

## Core details

<b>Asset ID</b>	MT-61	<b>Department</b>	Neonatology
<b>Device type</b>	Oxygen Concentrator	<b>Location</b>	Neonatology
<b>Manufacturer</b>	Medicor Elektronika	<b>Model</b>	JAY-10
<b>Serial number</b>	MZJ10D118316	<b>Year</b>	2021
<b>Responsible</b>	—	<b>Ownership</b>	Purchase

## Technical

<b>Power</b>	230V 880VA	<b>Connectivity</b>	—
<b>Other</b>	—		
<b>Software</b>	—		
<b>Consumables</b>	—		

## Maintenance log

PM = Preventive maintenance, CM = Corrective maintenance

Date	Type	Action	By	Ref/Ticket	Next due
—	PM	—	—	—	—

## Notes

Known issues

—

Spare parts on-site

—

Vendor / technician contact

—

Escalation



# Oxygen Concentrator

Out of service

Last updated 2026-05-21

Current location Neonatology

## Core details

<b>Asset ID</b>	MT-61	<b>Department</b>	Neonatology
<b>Device type</b>	Oxygen Concentrator	<b>Location</b>	Neonatology
<b>Manufacturer</b>	Philips	<b>Model</b>	EverFlo OPI
<b>Serial number</b>	2183670	<b>Year</b>	n/a
<b>Responsible</b>	—	<b>Ownership</b>	Purchase

## Technical

<b>Power</b>	n/a	<b>Connectivity</b>	—
<b>Other</b>	—		
<b>Software</b>	—		
<b>Consumables</b>	—		

## Maintenance log

PM = Preventive maintenance, CM = Corrective maintenance

Date	Type	Action	By	Ref/Ticket	Next due
—	PM	—	—	—	—

## Notes

Known issues

—

Spare parts on-site

—

Vendor / technician contact

—

Escalation



# Oxygen Concentrator

Operational

Last updated 2026-05-21

Current location Neonatology

## Core details

<b>Asset ID</b>	MT-62	<b>Department</b>	Neonatology
<b>Device type</b>	Oxygen Concentrator	<b>Location</b>	Neonatology
<b>Manufacturer</b>	NIDEK1	<b>Model</b>	MARK5 Nuvo 8 / OCSI
<b>Serial number</b>	20215862	<b>Year</b>	2020
<b>Responsible</b>	—	<b>Ownership</b>	Purchase

## Technical

<b>Power</b>	230V 490/625W	<b>Connectivity</b>	—
<b>Other</b>	—		
<b>Software</b>	—		
<b>Consumables</b>	—		

## Maintenance log

PM = Preventive maintenance, CM = Corrective maintenance

Date	Type	Action	By	Ref/Ticket	Next due
2025-09-26	PM	Maintenance	EK	—	2025-12-26

## Notes

Known issues

—

Spare parts on-site

—

Vendor / technician contact

—

Escalation



# Oxygen Concentrator

Out of service

Last updated 2026-05-21

Current location Neonatology

## Core details

<b>Asset ID</b>	MT-63	<b>Department</b>	Neonatology
<b>Device type</b>	Oxygen Concentrator	<b>Location</b>	Neonatology
<b>Manufacturer</b>	NIDEK1	<b>Model</b>	MARK5 Nuvo 8 / OCSI
<b>Serial number</b>	20215862	<b>Year</b>	2020
<b>Responsible</b>	—	<b>Ownership</b>	Purchase

## Technical

<b>Power</b>	230V 490/625W	<b>Connectivity</b>	—
<b>Other</b>	—		
<b>Software</b>	—		
<b>Consumables</b>	—		

## Maintenance log

PM = Preventive maintenance, CM = Corrective maintenance

Date	Type	Action	By	Ref/Ticket	Next due
— no entries yet —					

## Notes

Known issues

—

Spare parts on-site

—

Vendor / technician contact

—

Escalation



# Infusion Pump

Operational

Last updated 2026-05-21

Current location Neonatology

## Core details

<b>Asset ID</b>	MT-64	<b>Department</b>	Neonatology
<b>Device type</b>	Infusion Pump	<b>Location</b>	Neonatology
<b>Manufacturer</b>	GE	<b>Model</b>	n/a
<b>Serial number</b>	SH611110680SA	<b>Year</b>	2011
<b>Responsible</b>	—	<b>Ownership</b>	Purchase

## Technical

<b>Power</b>	n/a	<b>Connectivity</b>	—
<b>Other</b>	—		
<b>Software</b>	—		
<b>Consumables</b>	—		

## Maintenance log

PM = Preventive maintenance, CM = Corrective maintenance

Date	Type	Action	By	Ref/Ticket	Next due
— no entries yet —					

## Notes

Known issues

—

Spare parts on-site

—

Vendor / technician contact

—

Escalation



# CPAP Maschine

Operational

Last updated 2026-05-21

Current location Neonatology

## Core details

<b>Asset ID</b>	MT-65	<b>Department</b>	Neonatology
<b>Device type</b>	CPAP Maschine	<b>Location</b>	Neonatology
<b>Manufacturer</b>	Drive	<b>Model</b>	n/a
<b>Serial number</b>	A20A210054US	<b>Year</b>	n/a
<b>Responsible</b>	—	<b>Ownership</b>	—

## Technical

<b>Power</b>	n/a	<b>Connectivity</b>	—
<b>Other</b>	—		
<b>Software</b>	—		
<b>Consumables</b>	—		

## Maintenance log

PM = Preventive maintenance, CM = Corrective maintenance

Date	Type	Action	By	Ref/Ticket	Next due
2025-02-21	PM	Maintenance	EK	—	2025-09-21

## Notes

Known issues

—

Spare parts on-site

—

Vendor / technician contact

—

Escalation



# Electric Suction

Operational

Last updated 2026-05-21

Current location Neonatology

## Core details

<b>Asset ID</b>	MT-65	<b>Department</b>	Neonatology
<b>Device type</b>	Electric Suction	<b>Location</b>	Neonatology
<b>Manufacturer</b>	Smau	<b>Model</b>	YX932D
<b>Serial number</b>	A20A210054US	<b>Year</b>	2015
<b>Responsible</b>	—	<b>Ownership</b>	Purchase

## Technical

<b>Power</b>	220V 150VA	<b>Connectivity</b>	—
<b>Other</b>	—		
<b>Software</b>	—		
<b>Consumables</b>	—		

## Maintenance log

PM = Preventive maintenance, CM = Corrective maintenance

Date	Type	Action	By	Ref/Ticket	Next due
— no entries yet —					

## Notes

Known issues

—

Spare parts on-site

—

Vendor / technician contact

—

Escalation

